



Küppers Elektromechanik GmbH
Durchflussmessgeräte • Flow Meters

Liebigstraße 2 • D-85757 Karlsfeld • Tel. 0 81 31/5 93 91-0 • Fax 0 81 31/9 26 04 + 5 88 70

Certification

for the use of non-electrical equipment in hazardous areas.

KEM
Küppers Elektromechanik GmbH
Liebigstr. 2

D-85757 Karlsfeld

declare on their own responsibility that EC directive 94/9/EC does not apply to the mechanical turbine flow meters of the following series:

HM xxxxxxx

An analysis of the potential danger of causing an ignition proved that the equipment includes no potential flash source. Following the instructions stated in the associated product information the equipment may be used inside zone 1 and 2 for gas, haze and vapours.

This confirmation is based on an analysis of representative test samples of the series mentioned above.

Karlsfeld, 21.11.2003

Dipl.-Ing. Thomas Steuer
President

Franz Schmid
Ex protection representative